



# Site 06-053-1003: Salinas, CA

## Met One BAM-1020 Mass Monitor w/VSCC

Continuous (y) vs. FRM (x)  $PM_{2.5}$  ( $\mu g \cdot m^{-3}$ )




FRM minus Continuous  $PM_{2.5}$  ( $\mu g \cdot m^{-3}$ )




**Mean  $PM_{2.5}$  ( $\mu g \cdot m^{-3}$ )**

	#	FRM	Con.	Ratio (Con/FRM)
○	30	4.1	5.5	1.34
▽	13	3.8	4.9	1.29
△	1	1.8	2.2	1.22
▽	16	4.5	6.2	1.38
○	30	4.1	5.5	1.34

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData    △ Autumn    ▽ Spring    △ Summer    ▽ Winter  
8 2008 9 2009 0 2010