



Site 53-053-0029: Tacoma, WA

Thermo Scientific 8500C FDMS w/VSCC

Continuous (y) vs. FRM (x) $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




FRM minus Continuous $PM_{2.5}$ ($\mu g \cdot m^{-3}$)




Mean $PM_{2.5}$ ($\mu g \cdot m^{-3}$)

	#	FRM	Con.	Ratio (Con/FRM)
○	140	6.3	7	1.11
▽	59	9.6	10.3	1.07
△	10	3.6	3.9	1.08
▽	71	4	4.7	1.18
○	140	6.3	7	1.11

Additive (y) vs. Multiplicative (x) Bias



R (y) vs. FRM CCV (x)



○ AllData △ Autumn ▽ Spring △ Summer ▽ Winter
8 2008 9 2009 0 2010