

- 1. There is a 10% quarterly RAA requirement. If the normal concentration being measured is less than 10 ppm, or between 10 and 100 ppm, should they get the same sliding acceptability scale as is being used during the RATA?**

Yes, the quarterly RAA requirement is a sliding acceptability scale just like the RATA detailed in Section 13.1 of PS-16.