

**STL**

STL Knoxville
5815 Middlebrook Pike
Knoxville, TN 37921

Tel: 865 291 3000 Fax: 865 584 4315
www.stl-inc.com

ANALYTICAL REPORT

PROJECT NO. 142541

Focus/US Filter Westates TCO/G

Lot #: H6D030231

William Anderson

STL Knoxville
5815 Middlebrook Pike
Knoxville, TN 37921-5947

SEVERN TRENT LABORATORIES, INC.

A handwritten signature in black ink, appearing to read "K. S. Woodcock".

Kevin S. Woodcock
Project Manager

April 25, 2006

ANALYTICAL METHODS SUMMARY

H6D030231

| <u>PARAMETER</u> | <u>ANALYTICAL METHOD</u> |
|--|------------------------------|
| Gravimetric Analysis (GRA V) | EPA-18 GRAV |
| Total Chromatographable O rganics (TCO) Analysis | EPA-18 TCO |

References:

EPA-18 "GUIDANCE FOR TOTAL ORGANICS," EPA/600/R-96/033, MARCH 1996.

SAMPLE SUMMARY

H6D030231

| WO # | SAMPLE# | CLIENT | SAMPLE ID | SAMPLED DATE | SAMP TIME |
|-------|---------|--------------------|------------------------------|-----------------|--------------|
| H2H0E | 001 | G-2931/2932-R1-MM5 | FRONT HALF COMPOSITE TRAIN C | 03/28/06 | |
| H2H0G | 002 | G-2933/2934-R1-MM5 | BACK HALF COMPOSITE TRAIN C | 03/28/06 | |
| H2H0H | 003 | G-2935/2936-R1-MM5 | IMPINGER COMPOSITE TRAIN C | 03/28/06 | |
| H2H0J | 004 | G-3043/3044-R2-MM5 | FRONT HALF COMPOSITE TRAIN C | 03/29/06 | |
| H2H0M | 005 | G-3045/3046-R2-MM5 | BACK HALF COMPOSITE TRAIN C | 03/29/06 | |
| H2H0N | 006 | G-3047/3048-R2-MM5 | IMPINGER COMPOSITE TRAIN C | 03/29/06 | |
| H2H0R | 007 | G-3115/3116-R3-MM5 | FRONT HALF COMPOSITE TRAIN C | 03/30/06 | |
| H2H0W | 008 | G-3117/3118-R3-MM5 | BACK HALF COMPOSITE TRAIN C | 03/30/06 | |
| H2H0X | 009 | G-3119/3120-R3-MM5 | IMPINGER COMPOSITE TRAIN C | 03/30/06 | |
| H2H00 | 010 | G-3121/3122-R3-MM5 | FRONT HALF COMPOSITE BT C | 03/29/06 | |
| H2H01 | 011 | G-3123/3124-R3-MM5 | BACK HALF COMPOSITE BT C | 03/29/06 | |
| H2H02 | 012 | G-3125/3126-R3-MM5 | IMPINGER COMPOSITE BT C | 03/29/06 | |
| H2H03 | 013 | G-3127-R3-MM5 | TRAIN C XAD-2 TRIP/RB | 03/30/06 | |
| H2H05 | 014 | A-5380 | MEDIA CHECK XAD | 03/28/06 | |
| H2H06 | 015 | A-5382 | MEDIA CHECK FILTER | 03/28/06 | |

NOTE(S) :

- The analytical results of the samples listed above are presented on the following pages.
- All calculations are performed before rounding to avoid round-off errors in calculated results.
- Results noted as "ND" were not detected at or above the stated limit.
- This report must not be reproduced, except in full, without the written approval of the laboratory.
- Results for the following parameters are never reported on a dry weight basis: color, corrosivity, density, flashpoint, ignitability, layers, odor, paint filter test, pH, porosity pressure, reactivity, redox potential, specific gravity, spot tests, solids, solubility, temperature, viscosity, and weight.

PROJECT NARRATIVE

H6D030231

The results reported herein are applicable to the samples submitted for analysis only.

This report shall not be reproduced except in full, without the written approval of the laboratory.

The original chain of custody documentation is included with this report.

Sample Receipt

Custody seals were not present upon sample receipt at STL Knoxville; however, the samples were hand delivered.

The "Relinquished by" field on the chain of custody documentation did not contain a signature.

Quality Control

Unless otherwise noted, all holding times and QC criteria were met and the test results shown in this report meet all applicable NELAC requirements.

TCO/GRAV Sampling Train Preparation and Analysis

The semi-volatile organic sampling train components were extracted and analyzed for total chromatographable organics (TCO) and total gravimetric organics (GRAV) using STL Knoxville standard operating procedures KNOX-OP-0009 and KNOX-GC-0010, based on the following methods:

- SW-846 3542, "Extraction of Semivolatile Analytes Collected Using Method 0010 (Modified Method 5 Sampling Train)"
- SW-846 8015B, "Nonhalogenated Organics Using GC/FID"
- U.S. EPA "Guidance for Total Organics" (1996)

The sampling trains are prepared as three analytical fractions: The particulate filter and front half of the filter holder, nozzle and probe solvent rinses are combined as one sample. The XAD-2 resin trap and back half of the filter holder, coil condenser and

STL Knoxville maintains the following certifications, approvals and accreditations: Arkansas DEQ Cert. #05-043-0, California DHS ELAP Cert. #2423, Colorado DPHE, Connecticut DPH Cert. #PH-0223, Florida DOH Cert. #E87177, Georgia DNR Cert. #906 (SDWA, expires 6/24/05), Hawaii DOH, Illinois EPA Cert. #000687, Indiana DOH Cert. #C-TN-02, Iowa DNR Cert. #375, Kansas DHE Cert. #E-10349, Kentucky DEP Lab ID #90101, Louisiana DEQ Cert. #03079, Louisiana DOHH Cert. #LA030024, Maryland DHMH Cert. #277, Massachusetts DEP Cert. #M-TN009, Michigan DEQ Lab ID #9933, New Jersey DEP Cert. #TN001, New York DOH Lab #10781, North Carolina DPH Lab ID #21705, North Carolina DEHNR Cert. #64, Ohio EPA VAP Cert. #CL0059, Oklahoma DEQ ID #9415, Pennsylvania DEP Cert. #68-00576, South Carolina DHEC Lab ID #84001001, Tennessee DOH Lab ID #02014, Utah DOH Cert. #QUAN3, Virginia DGS Lab ID #00165, Washington DOE Lab #C120, West Virginia DEP Cert. #345, Wisconsin DNR Lab ID #998044300, US Army Corps of Engineers, Naval Facilities Engineering Service Center and USDA Soil Permit #S-46424. This list of approvals is subject to change and does not imply that laboratory certification is available for all parameters reported in this environmental sample data report.

PROJECT NARRATIVE

H6D030231

connecting glassware solvent rinses are combined as a separate sample. The condensate, impinger contents and their related glassware solvent rinses make up a third sample.

The filters and XAD components are spiked with the TCO surrogates and the components are Soxhlet extracted with methylene chloride. The condensates are spiked with the surrogates and extracted using a continuous liquid-liquid extractor. The extracts are concentrated to 10 mL and split in half with one portion for TCO analysis and the other for GRAV analysis.

Gas chromatography is used to determine the concentration of total chromatographable organics in the extracts. Detection, identification and quantification of the semi-volatile organic target analytes is done using a gas chromatograph with a flame ionization detector (GC/FID). This technique is non-specific in that all co-extracted organic compounds that elute between n-heptane and n-heptadecane, and render a flame ionization detector response, are detected. This range of elution approximately corresponds to the boiling range of 100°C to 300°C. The surrogate compound, n-heptadecane, is used to quantify sample recovery. Another surrogate compound, n-heptane, is also added to every sample, and the two surrogates are used to define the retention time window for TCO.

Total gravimetric organics are determined by evaporating a filtered portion of the sample extract at room temperature and determining the residue weight by difference. Residues having a boiling point of approximately 300°C and higher are determined by this procedure.

The preparation and analysis of method blanks is used to evaluate background response. Results are blank corrected by subtraction of the laboratory method blank result. The preparation and analysis of laboratory control samples are used to monitor test method performance. Gravimetric audit samples provide additional measurement of accuracy and precision for the gravimetric part of the test method.

TCO extract concentrations are calculated from peak areas as compared to a five point calibration curve. Sample results are calculated using the following equations:

$$\text{TCO Result, mg} = (\text{Extract concentration, mg / L}) * \text{Volume final extract, L} * \left(\frac{\text{Total Sample Volume, L}}{\text{Volume Sample Extracted, L}} \right)$$

STL Knoxville maintains the following certifications, approvals and accreditations: Arkansas DEQ Cert. #05-043-0, California DHS ELAP Cert. #2423, Colorado DPHE, Connecticut DPH Cert. #PH-0223, Florida DOH Cert. #E87177, Georgia DNR Cert. #906 (SDWA, expires 6/24/05), Hawaii DOH, Illinois EPA Cert. #000687, Indiana DOH Cert. #C-TN-02, Iowa DNR Cert. #375, Kansas DHE Cert. #E-10349, Kentucky DEP Lab ID #90101, Louisiana DEQ Cert. #03079, Louisiana DOHH Cert. #LA030024, Maryland DHMH Cert. #277, Massachusetts DEP Cert. #M-TN009, Michigan DEQ Lab ID #9933, New Jersey DEP Cert. #TN001, New York DOH Lab #10781, North Carolina DPH Lab ID #21705, North Carolina DEHNR Cert. #64, Ohio EPA VAP Cert. #CL0059, Oklahoma DEQ ID #9415, Pennsylvania DEP Cert. #68-00576, South Carolina DHEC Lab ID #84001001, Tennessee DOH Lab ID #02014, Utah DOH Cert. # QUAN3, Virginia DGS Lab ID #00165, Washington DOE Lab #C120, West Virginia DEP Cert. #345, Wisconsin DNR Lab ID #998044300, US Army Corps of Engineers, Naval Facilities Engineering Service Center and USDA Soil Permit #S-46424. This list of approvals is subject to change and does not imply that laboratory certification is available for all parameters reported in this environmental sample data report.

PROJECT NARRATIVE

H6D030231

$$\text{GRAV Result, mg} = (\text{Average Weight Sample and Beaker, mg} - \text{Beaker Tare Weight, mg}) * \text{SF} * \left(\frac{\text{Total Sample Volume, L}}{\text{Volume Sample Extracted, L}} \right)$$

Where: $\text{SF} = \text{Split Factor} = \left(\frac{\text{Volume Final Extract, mL}}{\text{Volume GRAV Extract, mL}} \right)$, which is typically 10 mL/5 mL.

Note: The term $\left(\frac{\text{Total Sample Volume, L}}{\text{Volume Sample Extracted, L}} \right) = 1$ for all non-aqueous samples and is used

when the volume of condensate exceeds one liter.

All sample results for TCO and GRAV are blank corrected using the following equation:

$$\text{Final Result, mg} = (\text{Sample Result, mg} - \text{Method Blank Result, mg})$$

A dilution factor was applied to the condensate portion of the analysis due to the initial volume of sample extracted. Only a portion of the total volume of condensate was used for this analysis. The dilution was used to properly adjust the results and reporting limits for the total amount of sample.

During review of the data, it appears that some of the Run 3 and Run 3 Blank Train samples may have been switched. The largest discrepancies appear for the condensate and filter fractions of the TCO analysis. However, this was not consistent with the GRAV analysis because the values were much lower.

STL Knoxville maintains the following certifications, approvals and accreditations: Arkansas DEQ Cert. #05-043-0, California DHS ELAP Cert. #2423, Colorado DPHE, Connecticut DPH Cert. #PH-0223, Florida DOH Cert. #E87177, Georgia DNR Cert. #906 (SDWA, expires 6/24/05), Hawaii DOH, Illinois EPA Cert. #000687, Indiana DOH Cert. #C-TN-02, Iowa DNR Cert. #375, Kansas DHE Cert. #E-10349, Kentucky DEP Lab ID #90101, Louisiana DEQ Cert. #03079, Louisiana DOHH Cert. #LA030024, Maryland DHMH Cert. #277, Massachusetts DEP Cert. #M-TN009, Michigan DEQ Lab ID #9933, New Jersey DEP Cert. #TN001, New York DOH Lab #10781, North Carolina DPH Lab ID #21705, North Carolina DEHNR Cert. #64, Ohio EPA VAP Cert. #CL0059, Oklahoma DEQ ID #9415, Pennsylvania DEP Cert. #68-00576, South Carolina DHEC Lab ID #84001001, Tennessee DOH Lab ID #02014, Utah DOH Cert. # QUAN3, Virginia DGS Lab ID #00165, Washington DOE Lab #C120, West Virginia DEP Cert. #345, Wisconsin DNR Lab ID #998044300, US Army Corps of Engineers, Naval Facilities Engineering Service Center and USDA Soil Permit #S-46424. This list of approvals is subject to change and does not imply that laboratory certification is available for all parameters reported in this environmental sample data report.

Sample Data Summary

STL Knoxville - ACS

Client Sample ID: G-2931/2932-R1-MM5 FRONT HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #...: H6D030231-001 Work Order #...: H2H0E1AA Matrix.....: AIR
 Date Sampled...: 03/28/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #...: 6095031
 Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.34 | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 89 | (50 - 150) |

STL Knoxville - ACS

Client Sample ID: G-2933/2934-R1-MM5 BACK HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-002 Work Order #....: H2H0G1AA Matrix.....: AIR
 Date Sampled....: 03/28/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095032
 Dilution Factor: 10 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 4.0 B | 0.50 | mg | 0.050 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | NC, DIL | (50 - 150) |

NOTE(S) :

NC The recovery and/or RPD were not calculated.

DIL The concentration is estimated or not reported due to dilution or the presence of interfering analytes.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-2935/2936-R1-MM5 IMPINGER COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-003 Work Order #....: H2H0H1AA Matrix.....: AIR
Date Sampled....: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
Prep Batch #....: 6095033
Dilution Factor: 2.53 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.98 B | 0.13 | mg | 0.018 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 92 | (50 - 150) |

NOTE (S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3043/3044-R2-MM5 FRONT HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-004 Work Order #....: H2H0J1AA Matrix.....: AIR
Date Sampled....: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
Prep Batch #....: 6095031
Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING</u> <u>LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------------|--------------|------------|
| Total Chromatographable Organics | 0.23 | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT</u> <u>RECOVERY</u> | <u>RECOVERY</u> <u>LIMITS</u> |
|------------------|-----------------------------------|----------------------------------|
| n-Heptadecane | 81 | (50 - 150) |

STL Knoxville - ACS

Client Sample ID: G-3045/3046-R2-MM5 BACK HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-005 Work Order #....: H2H0M1AA Matrix.....: AIR
Date Sampled....: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
Prep Batch #....: 6095032
Dilution Factor: 5 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 1.4 B | 0.25 | mg | 0.025 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 86 | (50 - 150) |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3047/3048-R2-MM5 IMPINGER COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-006 Work Order #....: H2H0N1AA Matrix.....: AIR
 Date Sampled...: 03/29/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095033
 Dilution Factor: 2.24 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 1.2 B | 0.11 | mg | 0.016 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 107 | (50 - 150) |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3115/3116-R3-MM5 FRONT HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-007 Work Order #....: H2H0R1AA Matrix.....: AIR
 Date Sampled....: 03/30/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095031
 Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.077 | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 87 | (50 - 150) |

STL Knoxville - ACS

Client Sample ID: G-3117/3118-R3-MM5 BACK HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-008 Work Order #....: H2H0W1AA Matrix.....: AIR
 Date Sampled....: 03/30/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095032
 Dilution Factor: 3 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 1.8 B | 0.15 | mg | 0.015 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 101 | (50 - 150) |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3119/3120-R3-MM5 IMPINGER COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-009 Work Order #....: H2H0X1AA Matrix.....: AIR
 Date Sampled....: 03/30/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095033
 Dilution Factor: 2.32 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.047 J,B | 0.12 | mg | 0.016 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 105 | (50 - 150) |

NOTE(S) :

J Estimated result. Result is less than RL.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3121/3122-R3-MM5 FRONT HALF COMPOSITE BT C

GC Semivolatiles

Lot-Sample #....: H6D030231-010 Work Order #....: H2H001AA Matrix.....: AIR
 Date Sampled...: 03/29/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095031
 Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.22 | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 89 | (50 - 150) |

STL Knoxville - ACS

Client Sample ID: G-3123/3124-R3-MM5 BACK HALF COMPOSITE BT C

GC Semivolatiles

Lot-Sample #....: H6D030231-011 Work Order #....: H2H011AA Matrix.....: AIR
 Date Sampled....: 03/29/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095032
 Dilution Factor: 3 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 1.1 B | 0.15 | mg | 0.015 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 93 | (50 - 150) |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3125/3126-R3-MM5 IMPINGER COMPOSITE BT C

GC Semivolatiles

Lot-Sample #....: H6D030231-012 Work Order #....: H2H021AA Matrix.....: AIR
 Date Sampled...: 03/29/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095033
 Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.49 B | 0.050 | mg | 0.0070 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 99 | (50 - 150) |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3127-R3-MM5 TRAIN C XAD-2 TRIP/RB

GC Semivolatiles

Lot-Sample #....: H6D030231-013 Work Order #....: H2H031AA Matrix.....: AIR
Date Sampled....: 03/30/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
Prep Batch #....: 6095032
Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.15 B | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 104 | (50 - 150) |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: A-5380 MEDIA CHECK XAD

GC Semivolatiles

Lot-Sample #....: H6D030231-014 Work Order #....: H2H051AA Matrix.....: AIR
Date Sampled....: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
Prep Batch #....: 6095032
Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING</u> <u>LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------------|--------------|------------|
| Total Chromatographable Organics | ND | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT</u> <u>RECOVERY</u> | <u>RECOVERY</u> <u>LIMITS</u> |
|------------------|-----------------------------------|----------------------------------|
| n-Heptadecane | 106 | (50 - 150) |

STL Knoxville - ACS

Client Sample ID: A-5382 MEDIA CHECK FILTER

GC Semivolatiles

Lot-Sample #....: H6D030231-015 Work Order #....: H2H061AA Matrix.....: AIR
 Date Sampled...: 03/28/06 Date Received...: 04/02/06
 Prep Date.....: 04/05/06 Analysis Date...: 04/16/06
 Prep Batch #....: 6095031
 Dilution Factor: 1 Method.....: EPA-18 TCO

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|-------------------------------------|---------------|----------------------------|--------------|------------|
| Total Chromatographable Organics | 0.0060 J | 0.050 | mg | 0.0050 |

| <u>SURROGATE</u> | <u>PERCENT RECOVERY</u> | <u>RECOVERY LIMITS</u> |
|------------------|-----------------------------|----------------------------|
| n-Heptadecane | 97 | (50 - 150) |

NOTE(S) :

J Estimated result. Result is less than RL.

STL Knoxville - ACS

Client Sample ID: G-2931/2932-R1-MM5 FRONT HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-001 Work Order #....: H2H0E1AC Matrix.....: AIR
Date Sampled....: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095034
Dilution Factor: 1 Method.....: EPA-18 GRAV

| PARAMETER | RESULT | REPORTING LIMIT | UNITS | MDL |
|----------------------------|----------|--------------------|-------|------|
| Total Gravimetric Organics | 0.47 J,B | 0.50 | mg | 0.18 |

NOTE(S):

J Estimated result. Result is less than RL.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-2933/2934-R1-MM5 BACK HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #...: H6D030231-002 Work Order #...: H2H0G1AC Matrix.....: AIR
Date Sampled...: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #...: 6095035
Dilution Factor: 1 Method.....: EPA-18 GRAV

| PARAMETER | RESULT | REPORTING LIMIT | UNITS | MDL |
|----------------------------|--------|--------------------|-------|------|
| Total Gravimetric Organics | 1.9 B | 0.50 | mg | 0.25 |

NOTE(S):

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-2935/2936-R1-MM5 IMPINGER COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #...: H6D030231-003 Work Order #...: H2H0H1AC Matrix.....: AIR
Date Sampled...: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #...: 6095036
Dilution Factor: 2.53 Method.....: EPA-18 GRAV

| PARAMETER | RESULT | REPORTING LIMIT | UNITS | MDL |
|----------------------------|--------|--------------------|-------|------|
| Total Gravimetric Organics | 0.68 J | 1.3 | mg | 0.63 |

NOTE(S) :

J Estimated result. Result is less than RL.

STL Knoxville - ACS

Client Sample ID: G-3043/3044-R2-MM5 FRONT HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #...: H6D030231-004 Work Order #...: H2H0J1AC Matrix.....: AIR
Date Sampled...: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #...: 6095034
Dilution Factor: 1 Method.....: EPA-18 GRAV

| PARAMETER | RESULT | REPORTING LIMIT | UNITS | MDL |
|----------------------------|----------|--------------------|-------|------|
| Total Gravimetric Organics | 0.40 J,B | 0.50 | mg | 0.18 |

NOTE(S) :

J Estimated result. Result is less than RL.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3045/3046-R2-MM5 BACK HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-005 Work Order #....: H2H0M1AC Matrix.....: AIR
Date Sampled....: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095035
Dilution Factor: 1 Method.....: EPA-18 GRAV

| PARAMETER | RESULT | REPORTING LIMIT | UNITS | MDL |
|----------------------------|--------|--------------------|-------|------|
| Total Gravimetric Organics | 1.3 B | 0.50 | mg | 0.25 |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3047/3048-R2-MM5 IMPINGER COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-006 Work Order #....: H2H0N1AC Matrix.....: AIR
Date Sampled...: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095036
Dilution Factor: 2.24 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | ND | 1.1 | mg | 0.56 |

STL Knoxville - ACS

Client Sample ID: G-3115/3116-R3-MM5 FRONT HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-007 Work Order #....: H2H0R1AC Matrix.....: AIR
Date Sampled....: 03/30/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095034
Dilution Factor: 1 Method.....: EPA-18 GRAV

| PARAMETER | RESULT | REPORTING LIMIT | UNITS | MDL |
|----------------------------|----------|--------------------|-------|------|
| Total Gravimetric Organics | 0.47 J,B | 0.50 | mg | 0.18 |

NOTE(S) :

J Estimated result. Result is less than RL.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3117/3118-R3-MM5 BACK HALF COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #...: H6D030231-008 Work Order #...: H2H0W1AC Matrix.....: AIR
Date Sampled...: 03/30/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #...: 6095035
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | 1.2 B | 0.50 | mg | 0.25 |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3119/3120-R3-MM5 IMPINGER COMPOSITE TRAIN C

GC Semivolatiles

Lot-Sample #....: H6D030231-009 Work Order #....: H2H0X1AC Matrix.....: AIR
Date Sampled....: 03/30/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095036
Dilution Factor: 2.32 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | ND | 1.2 | mg | 0.58 |

STL Knoxville - ACS

Client Sample ID: G-3121/3122-R3-MM5 FRONT HALF COMPOSITE BT C

GC Semivolatiles

Lot-Sample #....: H6D030231-010 Work Order #....: H2H001AC Matrix.....: AIR
Date Sampled....: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095034
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | 0.33 J,B | 0.50 | mg | 0.18 |

NOTE(S) :

J Estimated result. Result is less than RL.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3123/3124-R3-MM5 BACK HALF COMPOSITE BT C

GC Semivolatiles

Lot-Sample #....: H6D030231-011 Work Order #....: H2H011AC Matrix.....: AIR
Date Sampled....: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095035
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | 1.0 B | 0.50 | mg | 0.25 |

NOTE(S) :

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.

STL Knoxville - ACS

Client Sample ID: G-3125/3126-R3-MM5 IMPINGER COMPOSITE BT C

GC Semivolatiles

Lot-Sample #....: H6D030231-012 Work Order #....: H2H021AC Matrix.....: AIR
Date Sampled....: 03/29/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095036
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | ND | 0.50 | mg | 0.25 |

STL Knoxville - ACS

Client Sample ID: G-3127-R3-MM5 TRAIN C XAD-2 TRIP/RB

GC Semivolatiles

Lot-Sample #....: H6D030231-013 Work Order #....: H2H031AC Matrix.....: AIR
Date Sampled...: 03/30/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095035
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING</u> <u>LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------------|--------------|------------|
| Total Gravimetric Organics | ND | 0.50 | mg | 0.25 |

STL Knoxville - ACS

Client Sample ID: A-5380 MEDIA CHECK XAD

GC Semivolatiles

Lot-Sample #....: H6D030231-014 Work Order #....: H2H051AC Matrix.....: AIR
Date Sampled...: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095035
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | ND | 0.50 | mg | 0.25 |

STL Knoxville - ACS

Client Sample ID: A-5382 MEDIA CHECK FILTER

GC Semivolatiles

Lot-Sample #....: H6D030231-015 Work Order #....: H2H061AC Matrix.....: AIR
Date Sampled....: 03/28/06 Date Received...: 04/02/06
Prep Date.....: 04/05/06 Analysis Date...: 04/23/06
Prep Batch #....: 6095034
Dilution Factor: 1 Method.....: EPA-18 GRAV

| <u>PARAMETER</u> | <u>RESULT</u> | <u>REPORTING LIMIT</u> | <u>UNITS</u> | <u>MDL</u> |
|----------------------------|---------------|----------------------------|--------------|------------|
| Total Gravimetric Organics | 0.33 J,B | 0.50 | mg | 0.18 |

NOTE(S) :

J Estimated result. Result is less than RL.

B Method blank contamination. The associated method blank contains the target analyte at a reportable level.